


Author's Accepted Manuscript

Emerging Issues and Future Developments in Capsule Endoscopy

Piotr R. Slawinski, Keith L. Obstein, Pietro Valdastrì



www.techgiendoscopy.com/locate/

tgie

PII: S1096-2883(15)00009-1
DOI: <http://dx.doi.org/10.1016/j.tgie.2015.02.006>
Reference: YTGIE50449

To appear in: *Techniques in Gastrointestinal Endoscopy*

Cite this article as: Piotr R. Slawinski, Keith L. Obstein, Pietro Valdastrì, Emerging Issues and Future Developments in Capsule Endoscopy, *Techniques in Gastrointestinal Endoscopy*, <http://dx.doi.org/10.1016/j.tgie.2015.02.006>

This is a PDF file of an unedited manuscript that has been accepted for publication. As a service to our customers we are providing this early version of the manuscript. The manuscript will undergo copyediting, typesetting, and review of the resulting galley proof before it is published in its final citable form. Please note that during the production process errors may be discovered which could affect the content, and all legal disclaimers that apply to the journal pertain.

Emerging Issues and Future Developments in Capsule Endoscopy

Piotr R. Slawinski¹
Keith L. Obstein^{2,1}
Pietro Valdastrì^{1,2}

¹*STORM Lab, Department of Mechanical Engineering,
Vanderbilt University, Nashville, TN 37235-1592, USA*

²*Division of Gastroenterology, Hepatology, and Nutrition,
Vanderbilt University Medical Center, Nashville, TN 37235-1592, USA*

Corresponding author:

Piotr R. Slawinski

Address: STORM Lab, Department of Mechanical Engineering, Vanderbilt University
Olin Hall Room 406, 2400 Highland Avenue, Nashville, TN 37212

Telephone: 402-570-2864

Fax: none

Email: piotr.r.slawinski@vanderbilt.edu

The authors declare no conflict of interest.

Abstract

Capsule endoscopy (CE) has transformed from a research venture into a widely used clinical tool and the primary means for diagnosing small bowel pathology. These orally administered capsules traverse passively through the gastrointestinal tract via peristalsis and are used in the esophagus, stomach, small bowel, and colon. The primary focus of CE research in recent years has been enabling active CE manipulation and extension of the technology to therapeutic functionality; thus, widening the scope of the procedure. This review outlines clinical standards of the technology as well as recent advances in CE research. Clinical capsule applications are discussed with respect to each portion of the gastrointestinal tract. Promising research efforts are presented with an emphasis on enabling active capsule locomotion. The presented studies suggest, in particular, that the most viable solution for active capsule manipulation is actuation of a capsule via exterior permanent magnet held by a robot. Developing capsule procedures adhering to current healthcare standards, such as enabling a tool channel or irrigation in a therapeutic device, is a vital phase in the adaptation of CE in the clinical setting.

Highlights

- 1) Capsule endoscopy has become utilized in diagnosis of diseases of the esophagus, stomach, small intestine, and colon.
- 2) Current research is focused on active manipulation of capsule endoscopes that may enable both diagnostic and therapeutic interventions.
- 3) Manipulating capsule endoscopes via a robotically actuated magnet outside of the patient shows to be the most promising actuation technique to enable active maneuvering of the capsules.

Keywords

Medical robotics; Magnetic manipulation; Localization; Locomotion; Image enhancement

Abbreviations

CE – capsule endoscopy, CMOS – complementary metal oxide semiconductor, DoF – degree of freedom, FDA – food and drug administration, fps – frames per second, GERD – gastroesophageal reflux disease, GI – gastrointestinal, MAC – magnetic air capsule, MCE – magnetic capsule endoscopy, NBI – narrow band imaging, POSE – position and orientation, RF – radio frequency

Introduction

Since 1950 orally administered capsules with radio frequency transmission capability have been prototyped with an aim to study the physiological parameters of the gastrointestinal tract. A dearth of miniaturized electronic technology such as semiconductors and integrated circuits prevented development of these capsules until the beginning of the 21st century (1). Working independently, Gavriel Iddan (Israel) and Paul Swain (UK) introduced CE in 2000 as a means for providing patients with endoscopic imaging of the small bowel (2). First CE human trials were presented in 2001 by Given Imaging Ltd. (Yoqneam, Israel), who were the first to commercialize the technology (1). Given Imaging's first clinical CE marketed as M2A (Mouth-to-Anus) was awarded approval from the FDA in 2001. The second capsule to gain FDA approval was the M2A Plus, which was later remarketed under the now familiar name: PillCam®. Since the first FDA CE approval, over 2 million capsules have been ingested worldwide (3). The PillCam® series of capsules now encompasses about 95% of the CE market and has been used in over 1.7 million procedures worldwide and in more than 1,900 clinical studies (www.givenimaging.com). Today, leading CE companies include: Medtronic Inc. (USA) (Given Imaging Ltd. was acquired by Covidien Ltd. who was, in turn, acquired by Medtronic Inc. in 2014), Olympus Corporation (Japan), Chongqing Jinshan Science & Technology Co. Ltd, (China), IntroMedic Co. Ltd., (South Korea), and CapsoVision Inc. (USA). Originally developed for diagnostic use in the small bowel, CE application has spread to use in the esophagus, stomach, and colon. CE is still most widely used in the small bowel due to a lack of a noninvasive alternative. This review will examine current technology in clinical CE as well as the latest developments in image enhancement, investigation of active locomotion, and therapeutic possibilities.

Clinical Capsule Endoscopes

Esophageal Capsule Endoscopy

As opposed to the slow CE propagation through the small bowel, CE traversing the esophagus can reach speeds as high as 20 cm/s, leading to difficulties in assessing for pathology (4). Esophageal capsule primary indications include diagnosing reflux esophagitis, Barrett's esophagus, and varices (5). The most prevalent esophageal capsule, the PillCam® ESO (Given Imaging Ltd.), captures images at 7 fps from each of two cameras at its longitudinal ends, or a cumulative 14 fps. During a 2006 study of 42 patients with GERD, a PillCam ESO capturing images at 4 fps was compared to one capturing at 14 fps. The higher image capture rate was observed to be superior in visualization of the entire esophagus (76% vs. 12%, $P < 0.01$) making higher frame capture rate essential for esophageal CE (4). The diagnostic sensitivity and specificity reported for the ESO was 85.8% and 80.5% for varices, 98% and 100% for GERD, and 97% and 100% for Barrett's esophagus (6). Currently, the FDA-approved PillCam® ESO3 captures images at 35 fps compared to the ESO's 14 fps, has a field of view of 169°, and an operating time of no more than 30 minutes. CEs in the esophagus are well tolerated but remain limited by poor visualization, passive motion, and lack of therapeutic ability; and thus have not replaced flexible endoscopy for esophageal evaluation (5).

Small Bowel Capsule Endoscopy

CE is considered the gold standard for small bowel evaluation in patients with inflammatory bowel disease, suspected small bowel neoplastic lesions, and obscure GI bleeding (6–9). The PillCam® (Given Imaging Ltd.) is the most widely used capsule worldwide and the latest version, SB3 (FDA approved in 2013), acquires images at an adaptive rate of 2-6 fps.

The OMOM® Smart Capsule (Chongqing Jinshan Science and Technology Co., Chongqing, China) is the first capsule to include two-way data transmission via wearable RF sensors. This bidirectional communication allows for real time adjustment of image capture rate and light intensity

(10). The MiroCam® (IntroMedic Co., Ltd., South Korea) has similar parameters but is the only capsule to use electrical field propagation through the body – what is referred to as Human Body Communication (HBD) (www.intromedic.com). The EndoCapsule® (Olympus Corporation, Japan) operates similarly to the PillCam SB but uses high-resolution Charge Coupled Device technology for imaging rather than a CMOS sensor (www.olympusamerica.com) (6).

CapsoCam®SV-2 (CapsoVision® Inc., Saratoga, CA, USA) is the newest commercial CE and is the only one to employ on-board flash memory therefore requiring the patient to retrieve the capsule after passing it in their stool. Having undergone human trials (11), the CapsoCam takes a novel approach in providing a 360° panoramic view using 4 centralized cameras, each with >90° field of view. The CE has a battery life of 15 hours and captures images at 20 fps for the first 2 hours of operation and 12 fps for the duration of the procedure (12). The CapsoVision® system consists of the capsule, a stool retrieval tool: CapsoRetrieve®, the CapsoAccess® data retrieval system, and the CapsoView® application for viewing the images. Head to head trials with the aforementioned small bowel capsules have suggested comparable diagnostic yield, image quality, and completion rate (3). Various commercial capsules and related specifications are reported in Table 1.

Capsule Endoscopy in the Colon

Each year, colorectal cancer is the cause of nearly 608,000 deaths with approximately 14.5 million colonoscopies performed worldwide (13,14). As reported by www.cancer.org, nearly 100,000 cases of colon cancer are being diagnosed annually in the US alone while the number of occurrences is expected to increase by 62% by the year 2030 (15). Introducing a less invasive platform could prevent millions of patients from evading colonoscopy due to fear of an invasive procedure, procedural discomfort, bowel preparation, risk of adverse events, and the potential need for sedation (16,17). The most prevalent capsule for colonoscopy is the PillCam® COLON2 (Given Imaging Ltd.). The 31.5 mm × 11.6 mm capsule, which like the PillCam® ESO2, has a camera at both longitudinal ends and has capability of adjusting image capture rate from 4 fps in stasis, to 35 fps when in motion (6). In a study of

104 patients, the PillCam® COLON2's sensitivity and specificity for polyps ≥ 6 mm and ≥ 10 mm was 89% and 76%, and 88% and 89%, respectively (18). To accelerate CE transit through the colon and distend the lumen around the capsule, sodium phosphate has been added to the traditional pre-procedural patient solution of polyethylene glycol (19). The PillCam® COLON2 is actively used in Europe and Japan and has gained FDA clearance in 2014 in the United States for use in patients following an incomplete colonoscopy (www.givenimaging.com). Capsule colonoscopy is not yet considered a replacement to traditional colonoscopy due to specificity rates recorded as low as 64% (20). Impact due to lack of active steering techniques and inability to distend collapsed tissue may lead to reduced visibility (20).

Software

Apart from the capsule, CE systems generally include wearable sensors, data acquisition equipment, and video analysis software (6). About 55,000 images are captured during a typical lower-bowel CE procedure that are then post-processed (7). Clinicians are unable to observe these entire procedures in real time leading to development of CE image processing typically produced by the capsule manufacturer (i.e. Given Imaging's Rapid Reader, CapsoVisions's CapsoView, IntroMedic's MiroView and Olympus' VE-1 Real Time Viewer). The most widely used software of these is Given Imaging's Rapid Reader 8.0. Post-processing of data in software allows for filtering bandwidth to enhance surface contrast, make regions of interest more distinguishable, and apply algorithms to improve identification of blood. The most widely used tool in digital CE chromoendoscopy is Fujinon Intelligent Colon Enhancement (FICE®, Fujifilm, USA), which is used in the RAPID software and decomposes images into RGB wavelengths, increases contrast of each, and then reconstructs the image system (5,21,22).

Clinical CE systems have also incorporated a capsule localization feature in the wearable hardware. RF signal strength measured via wearable sensor allows for triangulation to determine CE position without the need for additional hardware. Though a seemingly simple solution to a convoluted problem of determining capsule position, using RF signal strength can, at best, be used to determine

global capsule coordinates, rather than locations with respect to anatomical landmarks. Further developments are needed to produce relevant localization data. Given Imaging has recently introduced the PillCam Recorder DR3 Simulator, an interactive data acquisition that readily acquires and analyzes capsule images during the procedure. This allows the image frame rate to be adjusted from 4 to 35 fps in accordance with capsule motion detected by image processing.

Current Research in Capsule Endoscopy

Image Enhanced CE

Image enhancement is used to improve standard colonoscopy procedures and is beginning to be applied in CE (22). Compared to healthy tissue, neoplastic intestinal tissue has been shown to exhibit a lower autofluorescent response to blue or ultraviolet light. A sarcoma-detecting capsule prototype has been developed that uses an autofluorescence intensity system. NBI and white light imaging would complement the system due to the high false-positive rate of autofluorescence intensity imaging. Further work is yet to be done to conduct in vivo trials (23). A method was introduced for gathering pathologic data from imaging. Tethered capsule endomicroscopy was introduced which utilizes optical frequency domain imaging technology in which images are obtained for single wavelength scattered light. Gora et al. introduced tethered capsule endomicroscopy, another imaging method utilizing optical frequency domain technology in which images are obtained for single wavelength scattered light. A tethered capsule produces 3D microstructural images of the esophagus while traversing passively via peristalsis and can be retracted after imaging is complete (24). This technology was first developed by NinePoint Medical Inc. under the name NvisionVLE™ (Cambridge, MA, USA) and FDA approved in 2013 for use in standard endoscopes but is not cleared for capsule use.

Capsule Localization

Accuracy in CE localization is vital for obtaining diagnostic information that can be used for planning treatment and for effectively actuating CEs via an external robot with a mounted magnet (25–

27). Localization can be referred to in two respects: relative to a point of interest inside the gastrointestinal tract, such as a lesion (diagnostic localization) or relative to a universal frame shared by an external driving robot (global localization). Due to the deformable nature of the GI tract, we cannot rely solely on global localization to locate target areas. Conversely, knowing only the distance from a capsule to a target area cannot provide enough information for control via an external robot. Though groups have reported 3D localization with an error of 1 mm using fluoroscopic imaging (28), developments must be made in describing a CE's location with respect to anatomical landmarks. The recently prototyped OdoCapsule targets this challenge by providing real-time distance from point of duodenal entry to exit from the ileum via mounted passive wheels on its side that operate as an "odometer". In vivo testing is pending to establish proof of this concept (29). In 2014, a hybrid localization technique utilizing video motion tracking and radio frequency was introduced by Bao et al. Experimental results show a localization error of less than 2.3 cm and a promising platform with no additional CE hardware (30). External magnet systems with hall-effect sensors have been studied for localization and pose (position and orientation) detection (26), demonstrating very promising results in two different in vivo studies on porcine models (31,32).

Capsule Locomotion

Upper GI and Stomach

As seen in table 1, nearly all commercially available CEs are limited to passive and random movement, hindering the technology from becoming a prevalent diagnostic technique throughout the entire GI tract (33). An upper GI CE actuation technique must allow for resistance of peristaltic force in the tubular esophagus, remain dexterous enough to directly observe the gastroesophageal junction by rotating the capsule body 180°, and maintain compliance while moving through the open fluid-filled gastric lumen, pylorus, and duodenum; resisting gastric motility. Tortuous design requirements like these have provoked a ubiquitous need for an active capsule actuation method. The varying geometry throughout the GI tract has resulted in a research focus on CE actuation using a magnet maneuvered on

the outside of a patient's body. Magnetic coupling, a means of exerting forces across a physical barrier, has been a focus of CE active locomotion research in the last 4 years and is one of the four "grand challenges" of CE, along with magnetic field gradient control, accurate localization, and enabling diagnostic and therapeutic functions (17,33). Applying force from outside the body enables a clinician to control the point of desired manipulation on a capsule, such as direct steering of the camera, versus pushing from the distal end of the endoscope. Benefits of magnetic capsule endoscopy (MCE) as outlined, has made MCE an American Society for Gastrointestinal Endoscopy emerging technology focus and subject of research worldwide (34).

Keller et al. performed the first human MCE studies in 2010. During the first study, a PillCam® COLON was modified to house a permanent magnet and was manipulated by a hand-held magnet in a volunteer's esophagus and stomach (35). Later that year, Keller et al. performed separate studies on MCE manipulation in both the esophagus and stomach; also actuated via hand-held magnet. During the esophageal study, PillCam® ESO2 capsules with embedded permanent magnets were swallowed by 10 healthy subjects and actuated by an external hand held magnet. Though the procedure was reported as safe and feasible, magnetic forces were not strong enough to resist esophageal peristalsis (36). During the latter study, CE maneuverability in 7 of 10 subjects in the stomach was referred to as excellent but visualization of the gastric mucosa was not complete as magnetic forces were not strong enough to overcome passage from the antrum to the duodenum during phase III motility (37). Since these feasibility trials, two MCE systems that utilize a hand-held external magnet have become commercially available: the OMOM Controllable Capsule System (Chongqing Jinshan Science & Technology, China) and the MiroCam™ Navi (IntroMedic Co., Ltd., South Korea). The authors have no knowledge of effectiveness of these systems at the time of writing this review.

Though hand held magnet control may allow for faster movement through a straight trajectory, the use of an external robot with a permanent magnet mounted to the end effector has been shown to be more precise when approaching a target point, which is imperative for diagnosis and therapeutic

maneuvering (38). One of the first groups to study external robotic control was Carpi et al. who first conducted proof-of-concept in vivo porcine studies (2009 into 2011). During this study, a PillCam capsule was actuated via the Niobe magnetic navigation system (Stereotaxis, Inc., St. Louis, MO) in a porcine model (28,39).

Human trials using external robotic actuation were presented in 2010 by Rey et al. who conducted a feasibility study on the magnetically guided capsule endoscope introduced by Siemens Healthcare and Olympus Medical Systems Corporation. The study included 53 subjects who were placed on a bed and positioned under a magnetic guidance system that produced gradients in 3 dimensional space around the subjects' water-distended stomachs. Capsules were guided magnetically using the MRI-like system and could be maneuvered at the gastric water surface or made to dive to the bottom of the gastric lumen. To improve capsule mobility, subjects were at times asked to rotate their body while lying on the bed suggesting a deficiency in the control method. The procedure lasted an average of 30 minutes with an overall technical success rate of 98% with full visualization of the antrum 98%, body 96%, fundus 73%, and cardia 75% (40). A similar study was conducted in 2012 where a 5-degree of freedom robot was utilized to actuate capsules. An assessment of the actuation system was conducted on 34 healthy volunteers who ingested magnetic capsules and powder for gastric distension. The authors concluded that examination of the human stomach via MCE is feasible and safe (41).

The use of magnetism for GI CE control has not been limited to manipulation by an external robot. Conversely to the aforementioned robot manipulation techniques, Kim et al. in 2014 studied the feasibility of patients wearing a belt with an array of 17 total magnets to prohibit a swallowed capsule from advancing past the stomach. A modified MiroCam (IntroMedic Co., Ltd., South Korea) capsule was equipped with internal permanent magnet disks and results suggest that real-time monitoring of gastric motility via MCE is feasible (42). Besides magnetic actuation in the stomach, investigators have studied feasibility of submarine-like capsule robots for enabling omnidirectional swimming through a fluid filled gastric lumen. This technology remains in preliminary stages of development (43,44).

Lower GI

Though the narrow diameter of the small intestine allows for adequate tissue visibility with CE, the larger diameter lumen of the colon tends to collapse, inhibiting visibility. Maneuverability through such collapsed tissue is necessary for inspecting behind folds and thus performing a complete exam. Since the advent of the technology, the main actuation modes of CEs can be categorized as onboard or external. Onboard actuation consists of miniaturized mechanical hardware built into the capsule allowing them to self-propel through the GI tract, while external actuation consists of pulling a capsule via magnetic coupling.

The most commonly studied onboard actuation techniques are legged locomotion and treaded driving. A number of legged capsule prototypes were developed in hopes of enabling crawling actuation in the small bowel. Mechanical complexity and convoluted control of such legged capsules has hindered further development and seems to lack feasibility for commercialization (45–47). In 2012, the tethered robotic capsule endoscope (tRCE) was developed which is driven via four series of treads on each of its sides. Designed for mobility in a collapsed lumen, the treads are made to maximize contact area with tissue thus increasing friction. Due to agility and virtually unlimited power supply, capsules like such can be incorporated into other areas of minimally invasive surgery (48–50). Onboard actuation techniques allow for accurate relative motion but are limited by high power consumption, increased capsule volume (needed for hardware), and general complexity of miniature mechanical systems.

Similar to the upper GI tract, research into active robotic CE driving began with passive movement of a permanent magnet mounted on a robotic arm. Early in vivo trials involving actuating a MCE through the lower GI tract via a passively actuated external magnet mounted on a hydraulic arm occurred as early as 2008 (51). Since then, robotic platforms have been developed in pursuit of intuitive closed loop CE control via an external permanent magnet in various quantities and configurations. Procedural accuracy of control using an external industrial robot for MCE has been studied and concluded to be feasible (52,53). In 2014, Sun et al. presented a magnetic driving system consisting of two external

magnets placed on the lateral and medial side of a patient to produce coplanar motion of an internal capsule with the magnets. The study involved magnetic and hybrid (legged) capsule actuation, where purely magnetic driving was concluded as best (54).

As opposed to the above techniques of pulling a CE via magnetic field gradient, an actuation approach utilizing rotating magnetic field generated by spinning an external permanent magnet has been developed. First introduced in 2003 by Olympus Optical Co. Ltd. (patent no. US20030181788 A1), these screw type capsules are rotated about their longitudinal axis. Thread-like features on the external capsule shell allow for translation of magnet-induced rotation to lateral capsule motion. Studies on spiral actuation were conducted in 2012 by equipping a PillCam® SB2 with external threads (55). As opposed to direct magnetic gradient dependence of aforementioned MCE guidance techniques, MCE actuation via rotating permanent magnet is governed by torque induced by magnetic field intensity that decays slower with distance, and can allow for more controlled actuation (55–57).

A locomotion technique utilizing both onboard and external actuation has been developed. Hybrid locomotion, defined by Simi et al. in (58) as a combination between internal actuation mechanisms and external magnetic dragging, can provide a way for CEs to anchor against peristalsis and make accurate movements without the need for on board elements to provide total drive power; thus reducing power requirements (59). One of the first hybrid designs was developed in 2009 and tested in vivo via porcine model that was actuated via external magnetic fields and, with a 3-leg mechanism, was able to distend collapsed tissue (58). Another hybrid system was proposed by Ciuti et al. who utilized vibration to reduce contact friction between a CE and tissue while the capsule was dragged magnetically (60).

Adverse events in colonoscopy generally arise due to the mechanics of pushing a flexible, yet stiff, colonoscope through a pliable lumen, where device buckling or colon looping can occur. Such events can be averted by pulling the device through the colon via external magnetic force (17).

Aforementioned studies show that MCE driving via external robot is a favorable capsule actuation

technique; however, research addressing tissue interaction must be performed. Researchers are challenged with designing MCEs to meet miniaturized size constraints, ability for active locomotion, and adequate power supply; parameters which are unfortunately antithetical. The most pragmatic solution is to introduce a tether with tool channel to enable standard instruments to be used under robotic manipulation of the MCE. Valdastrì et al introduced a tethered MCE system with tool channel in 2011. The Magnetic Air Capsule (MAC), pictured in figure 1, is steered via 6 degrees of freedom (DoF) industrial robot and contains a camera, illumination, and a tool channel as used in standard colonoscopes. During in vivo trials, biopsy samples were collected with the MAC that were of the same size as those collected during traditional colonoscopy (61). A 2012 MAC feasibility study including 22 subjects using a 7 DoF robot assessed navigation of an MCE through an ex vivo porcine colon where experts and trainees were tasked with detecting a total of 672 pins. Pin detection was 80.9% for CE and 85.8% for traditional colonoscopy with appropriate diagnostic yield. The authors concluded that the MCE was a feasible procedure (53). Real time localization must be applied to the system to enable closed loop control (26). Global localization will prevent loss of magnetic coupling between robot and MCE, while diagnostic localization will maintain physician awareness of lesion position. The act of pulling a small capsule with flexible endoscope will reduce risk of tissue stress, colon looping, and show pursuit of promise in replacing traditional colonoscopy (17).

Interaction with tissue

Studies on therapeutic capsules have been conducted which address wireless power transfer to CEs by induction, utilizing autofluorescence in diagnosing tissue, and various tissue biopsy techniques (23,62–66). In 2014, a novel capsule targeted for drug delivery in the GI tract was introduced. In vivo porcine studies were conducted on the drug delivery needle capsule targeted for delivery of biologic therapeutics that could otherwise be adversely affected by bacteria and pH levels in the GI tract. The capsule is covered with radially oriented needles embedded in a pH sensitive coating that both protect tissue while the capsule is swallowed and keeps drug contained. Once the capsule reaches the target

delivery site, the coating dissolves and the drug is squeezed out via peristalsis into hollow needles that penetrate tissue, and detach from the capsule, remaining embedded in tissue, allowing for steady drug release (67).

Designing CEs to achieve accurate CE locomotion and ability to resist or drive against bowel forces requires GI force characterization. Terry et al. were the first to quantify contractile strength of the porcine small intestine in vivo while studying force variability in the proximal, middle, and distal regions. Measuring the migrating motor complex force with sensor balloon segments, contact force values in the distal small bowel were observed to be 92% larger than in the proximal small bowel (68). Knowledge of force variations as above is vital for control algorithm development. In 2014, Di Natali et al. presented a real-time platform for quantifying the propelling force required for traversing the GI tract. The capsule, while coupled with an external permanent magnet, sent wireless data to an external transceiver consisting of inter-magnetic force and capsule acceleration, from which a force model was developed (32). Care must be taken in MCE to maintain streamlined capsule motion while preventing overstressing of tissue due to excessive magnetic link strength that may cause patient discomfort, tissue ischemia, or tissue perforation.

Conclusions and Future Direction

CE is the primary mode of small bowel examination and the technology has expanded to clinical applications in the esophagus, stomach, and colon. Current CEs are limited to acquiring images of the mucosa while being passively moved throughout the GI tract via peristalsis. The primary focus of CE research in recent years has been enabling active CE manipulation and extension of the technology to therapeutic functionality. The most promising mode of CE control involves actuating tethered capsules via external permanent magnets manipulated by a robotic controller. Actuation of magnetic or hybrid capsules will allow for more thorough bowel examination and the ability for therapeutic intervention. Attaining such control in the clinical setting will serve as a paradigm shift. Development of these control

algorithms may contribute to the advancement of CE's use in practice and possibly eliminate the need for patient sedation within the next 5-10 years.

CE development is a labor and resource intensive task involving determination of functional parameters, designing mechanical components, developing miniature circuitry, and integrating custom hardware and software into a functioning model. Work has recently been done to develop open source hardware and software for expediting CE development and a web-based collaborative design environment (69,70). Creating easy-to-use components may enable researchers to avoid the time consuming process of gaining expertise in various domains such as low miniaturized electronics or level programming and skip to design for creative application. Efforts like these, and the research ventures outlined in this review suggest vast advancements in CE technology in the coming 5 years. The future of CE is promising, and further clinical applications of CE actuation technology and therapeutic ability is expected.

Acknowledgements

Research reported in this publication was supported in part by the National Institute of Biomedical Imaging And Bioengineering of the National Institutes of Health under Award Number R01EB018992, and in part by the National Science Foundation under grant number CNS-1239355. This material is also based upon work supported by the National Science Foundation Graduate Research Fellowship Program under Grant No. 1445197. Any opinions, findings and conclusions or recommendations expressed in this material are those of the authors and do not necessarily reflect the views of the National Institutes of Health or the National Science Foundation.

Conflict of Interest Statement

The authors declare no conflict of interest.

References

1. Pan G, Wang L: Swallowable Wireless capsule endoscopy: progress and technical challenges. *Gastroenterol Res Pract* 2012:1-9, 2011
2. Iddan G, Meron G, Glukhovsky A, et al: Wireless capsule endoscopy. *Nature* 405: 417-418, 2000
3. Koulaouzidis A, Rondonotti E, Karargyris A: Small-bowel capsule endoscopy: a ten-point contemporary review. *World J Gastroenterol WJG*. 19:3726-3746, 2013
4. Koslowsky B, Jacob H, Eliakim R, Adler SN: PillCam ESO in esophageal studies: improved diagnostic yield of 14 frames per second (fps) compared with 4 fps. *Endoscopy*. 38(1):27–30, 2006
5. Hale MF, Sidhu R, McAlindon ME: Capsule endoscopy: Current practice and future directions. *World J Gastroenterol WJG*. 20:7752-7759, 2014

6. Goenka MK, Majumder S, Goenka U: Capsule endoscopy: Present status and future expectation. *World J Gastroenterol WJG*. 20:10024-10037, 2014
7. Mustafa BF, Samaan M, Langmead L, et al: Small bowel video capsule endoscopy: an overview. *Expert Rev Gastroenterol Hepatol*. 7(4):323–329, 2013
8. Rey J-F: The future of capsule endoscopy. *Keio J Med*. 62(2):41–46, 2013
9. Valdastrì P, Simi M, Webster RJ: Advanced Technologies for Gastrointestinal Endoscopy. *Annu Rev Biomed Eng*. 14:397–429, 2012
10. Mackiewicz M: State of the technology and computer vision tools after the first decade, new techniques in gastrointestinal endoscopy. *InTech*. 104-124, 2011. <http://www.intechopen.com/books/new-techniques-in-gastrointestinal-endoscopy/capsule-endoscopy-state-of-the-technology-and-computer-vision-tools-after-the-first-decade>
11. Tontini GE, Cavallaro F, Neumann H, et al: Extensive small-bowel Crohn's disease detected by the newly introduced 360° panoramic viewing capsule endoscopy system. *Endoscopy*. 46:353-354, 2014
12. Koulaouzidis A, Dabos KJ: Looking forwards: not necessarily the best in capsule endoscopy? *Ann Gastroenterol* 26: 365-367, 2013
13. Ferlay J, Shin H-R, Bray F, et al: Estimates of worldwide burden of cancer in 2008: GLOBOCAN 2008. *Int J Cancer J Int Cancer*. 127:2893-2917, 2010
14. Seeff LC, Richards TB, Shapiro JA, et al: How many endoscopies are performed for colorectal cancer screening? Results from CDC's survey of endoscopic capacity. *Gastroenterology*.127:1670-1677, 2004
15. O'Callaghan T: The prevention agenda. *Nature*. 471:S2-S4, 2011

16. Bujanda L, Sarasqueta C, Zubiaurre L, et al: Low adherence to colonoscopy in the screening of first-degree relatives of patients with colorectal cancer. *Gut*. 56:1714-1718, 2007
17. Obstein K, Valdastrì P: Advanced endoscopic technologies for colorectal cancer screening. *World J Gastroenterol WJG*. 19:431-439, 2013
18. Eliakim R, Yassin K, Niv Y, et al: Prospective multicenter performance evaluation of the second-generation colon capsule compared with colonoscopy. *Endoscopy*. 41:1026-1031, 2009
19. Spada C, De Vincentis F, Cesaro P, et al: Accuracy and safety of second-generation PillCam COLON capsule for colorectal polyp detection. *Ther Adv Gastroenterol*. 5:173-178, 2012
20. Riccioni ME, Urgesi R, Cianci R, et al: Colon capsule endoscopy: Advantages, limitations and expectations. Which novelties? *World J Gastrointest Endosc*. 4:99-107, 2012
21. Rey JW, Kiesslich R, Hoffman A: New aspects of modern endoscopy. *World J Gastrointest Endosc*. 6:334-344, 2014
22. Ibrahim M, Van Gossum A: Novel imaging enhancements in capsule endoscopy. *Gastroenterol Res Pract*. 2013:1-5, 2013
23. Al-Rawhani MA, Chitnis D, Beeley J, et al: Design and implementation of a wireless capsule suitable for autofluorescence intensity detection in biological tissues. *IEEE Trans Biomed Eng*. 60:55-62, 2013
24. Gora MJ, Sauk JS, Carruth RW, et al: Tethered capsule endomicroscopy enables less invasive imaging of gastrointestinal tract microstructure. *Nat Med*. 19(2):238-240, 2013
25. Pahlavan K, Bao G, Ye Y, et al: RF localization for wireless video capsule endoscopy. *Int J Wireless Inf Networks*. 19:326-340, 2012

26. Di Natali C, Beccani M, Valdastri P: Real-time pose detection for magnetic medical devices. *IEEE Trans Magn.* 49(7):3524–3527, 2013
27. Ciuti G, Valdastri P, Menciassi A, et al: Robotic magnetic steering and locomotion of capsule endoscope for diagnostic and surgical endoluminal procedures. *Robotica.* 28(2):199–207, 2010
28. Carpi F, Kastelein N, Talcott M, et al: Magnetically controllable gastrointestinal steering of video capsules. *IEEE Trans Biomed Eng.* 58(2):231–234, 2011
29. Karargyris A, Koulaouzidis A: OdoCapsule: Next generation wireless capsule endoscopy with accurate localization and video stabilization. *IEEE Trans Biomed Eng* (in press) DOI: 10.1109/TBME.2014.2352493
30. Bao G, Pahlavan K, Mi L: Hybrid localization of micro-robotic endoscopic capsule inside small intestine by data fusion of vision and RF sensors. *IEEE Sens J.* 2014:1530-1537, 2014
31. Beccani M, Di Natali C, Sliker LJ, et al: Wireless tissue palpation for intraoperative detection of lumps in the soft tissue. *IEEE Trans Biomed Eng.* 61(2):353–361, 2014
32. Natali CD, Beccani M, Obstein KL, et al: A wireless platform for in vivo measurement of resistance properties of the gastrointestinal tract. *Physiol Meas.* 35(7):1197–1214, 2014
33. Carpi F, Shaheed H: Grand challenges in magnetic capsule endoscopy. *Expert Rev Med Devices.* 10(4):433–436, 2013
34. Gottlieb KT, Banerjee S, Barth BA, et al: Magnets in the GI tract. *Gastrointest Endosc.* 72(4):561–567, 2013
35. Swain P, Toor A, Volke F, et al: Remote magnetic manipulation of a wireless capsule endoscope in the esophagus and stomach of humans (with videos). *Gastrointest Endosc.* 71(7):1290–1293, 2010

36. Keller J, Fibbe C, Volke F, et al: Remote magnetic control of a wireless capsule endoscope in the esophagus is safe and feasible: results of a randomized, clinical trial in healthy volunteers. *Gastrointest Endosc.* 72(5):941–946, 2010
37. Keller J, Fibbe C, Volke F, et al: Inspection of the human stomach using remote-controlled capsule endoscopy: a feasibility study in healthy volunteers (with videos). *Gastrointest Endosc.* 73:22–28, 2011
38. Ciuti G, Donlin R, Valdastrì P, et al: Robotic versus manual control in magnetic steering of an endoscopic capsule. *Endoscopy.* 42(2):148–152, 2010
39. Carpi F, Pappone C: Magnetic maneuvering of endoscopic capsules by means of a robotic navigation system. *IEEE Trans Biomed Eng.* 56(5):1482–1490, 2009
40. Rey JF, Ogata H, Hosoe N, et al: Feasibility of stomach exploration with a guided capsule endoscope. *Endoscopy.* 42(7):541–545, 2010
41. Liao Z, Duan X-D, Xin L, et al: Feasibility and safety of magnetic-controlled capsule endoscopy system in examination of human stomach: a pilot study in healthy volunteers. *J Interv Gastroenterol.* 2(4):155–160, 2012
42. Kim HM, Choi JS, Cho JH: A pilot trial of ambulatory monitoring of gastric motility using a modified magnetic capsule endoscope. *J Neurogastroenterol Motil.* 20(2):261–264, 2014
43. Tortora G, Valdastrì P, Susilo E, et al: Propeller-based wireless device for active capsular endoscopy in the gastric district. *Minimally Invasive Therapy.* 18:280-290, 2009
44. De Falco I, Tortora G, Dario P, et al: An integrated system for wireless capsule endoscopy in a liquid-distended stomach. *IEEE Trans Biomed Eng.* 61(3):794–804, 2014


45. Valdastri P, Webster RJ, Quaglia C, et al: A new mechanism for mesoscale legged locomotion in compliant tubular environments. *IEEE Trans Robot.* 25(5):1047–1057, 2009
46. Kim HM, Yang S, Kim J, et al: Active locomotion of a paddling-based capsule endoscope in an in vitro and in vivo experiment (with videos). *Gastrointest Endosc.* 72(2):381–387, 2010
47. Chen W, Yan G, Wang Z, et al: A wireless capsule robot with spiral legs for human intestine. *Int J Med Robot Comput Assist Surg MRCAS.* 10(2):147–161, 2014
48. Sliker LJ, Kern MD, Schoen JA, et al: Surgical evaluation of a novel tethered robotic capsule endoscope using micro-patterned treads. *Surg Endosc.* 26(10):2862–2869, 2012
49. Sliker LJ, Kern MD, Rentschler ME. An automated traction measurement platform and empirical model for evaluation of rolling micropatterned wheels. *IEEE/ASME Trans Mechatron.* DOI: 10.1109/TMECH.2014.2357037 (in press)
50. Kern MD, Alcaide JO, Rentschler ME: Soft material adhesion characterization for in vivo locomotion of robotic capsule endoscopes: Experimental and modeling results. *J Mech Behav Biomed Mater.* 39:257–269, 2014
51. Valdastri P, Quaglia C, Susilo E, et al: Wireless therapeutic endoscopic capsule: in vivo experiment. *Endoscopy.* 40(12):979–982, 2008
52. Ciuti G, Salerno M, Lucarini G, et al: A comparative evaluation of control interfaces for a robotic-aided endoscopic capsule platform. *IEEE Trans Robot.* 28(2):534–538, 2012
53. Arezzo A, Menciassi A, Valdastri P, et al: Experimental assessment of a novel robotically-driven endoscopic capsule compared to traditional colonoscopy. *Dig Liver Dis Off J Ital Soc Gastroenterol Ital Assoc Study Liver.* 45(8):657–662, 2013

54. Sun ZJ, Cheng XG, Cao S, et al: Multi-applications of a magnet configuration in actuating capsule endoscope. 2014 IEEE/ASME International Conference on Advanced Intelligent Mechatronics (AIM). 106–111, 2014
55. Zhou H, Alici G, Than TD, et al: Modeling and experimental investigation of rotational resistance of a spiral-type robotic capsule inside a real intestine. IEEEASME Trans Mechatron. 18(5):1555–1562, 2013
56. Lee JS, Kim B, Hong YS. A flexible chain-based screw propeller for capsule endoscopes. Int J Precis Eng Manuf. 10(4):27–34, 2009
57. Mahoney AW, Abbott JJ. Generating rotating magnetic fields with a single permanent magnet for propulsion of untethered magnetic devices in a lumen. IEEE Trans Robot. 30(2):411–420, 2014
58. Simi M, Valdastrì P, Quaglia C, et al: Design, fabrication, and testing of a capsule with hybrid locomotion for gastrointestinal tract exploration. IEEEASME Trans Mechatron. 15(2):170–180, 2010
59. Basar MR, Malek F, Juni KM, et al: Ingestible wireless capsule technology: A review of development and future indication. Int J Antennas Propag. 2012:1-14, 2012
60. Ciuti G, Pateromichelakis N, Sfakiotakis M, et al: A wireless module for vibratory motor control and inertial sensing in capsule endoscopy. Procedia Eng. 25:92–95, 2011
61. Valdastrì P, Ciuti G, Verbeni A, et al: Magnetic air capsule robotic system: proof of concept of a novel approach for painless colonoscopy. Surg Endosc. 26(5):1238–1246, 2012
62. Chen W, Yan G, He S, et al: Wireless powered capsule endoscopy for colon diagnosis and treatment. Physiol Meas. 34(11):1545–1561, 2013

63. Kong K, Yim S, Choi S, Jeon D: A robotic biopsy device for capsule endoscopy. *J Med Devices*. 6(3):031004-1–9, 2012
64. Yim S, Sitti M: Design and analysis of a magnetically actuated and compliant capsule endoscopic robot. 2011 IEEE International Conference on Robotics and Automation (ICRA). 4810–4815, 2011
65. Simi M, Gerboni G, Menciassi A, et al: Magnetic torsion spring mechanism for a wireless biopsy capsule. *J Med Devices*. 7(4):041009, 2013
66. Woods SP, Constandinou TG: Wireless capsule endoscope for targeted drug delivery: Mechanics and design considerations. *IEEE Trans Biomed Eng*. 60(4):945–953, 2013
67. Traverso G, Schoellhammer CM, Schroeder A, et al: Microneedles for Drug Delivery via the Gastrointestinal Tract. *J Pharm Sci. Rapid Comm*:1-6, 2014
68. Terry BS, Schoen JA, Rentschler ME: Measurements of the contact force from myenteric contractions on a solid bolus. *J Robot Surg*. 7:53–57, 2013
69. Beccani M, Susilo E, Di Natali C, et al: SMAC — A Modular Open Source Architecture for Medical Capsule Robots. *Int J Adv Robot Syst. Open Access*:1-16, 2014
70. Beccani, M, Tunc H, Taddese, et al: A. Systematic Design of Medical Capsule Robots. (in press)

Legends for Illustrations

Fig 1 – Magnetic air capsule system



Legends for Tables

Table 1 – GI capsules in clinical use today

Table 1.

Capsule	Purpose	Company	FDA	FPS	Size (mm)	Software	Related Devices
PillCam® ESO 3	Esophageal Imaging	Given Imaging Ltd. (Israel)	Yes	35	26.0 × 11.0	RAPID®v8.0	PillCam® Recorder DR3, PillCam® Express, Sensor Belt
PillCam® SB 3	Small Bowel (SB) Imaging		Yes	2 to 6	26.0 × 11.0		
PillCam® COLON 2	Colon Imaging		Yes	4 to 35	31.5 × 11.6		
PillCam® SmartPill	Sensory		Yes	No Imaging	26.0 × 13.0	N/A	N/A
Bravo® Capsule	Sensory		Yes	No Imaging	26.7 × 6.3 × 5.4	N/A	N/A
Agile™ Patency Capsule	Patency		Yes	No Imaging	26.4 × 11.4	None. Dissolvable capsule	None
CapsoCam® SV-2	SB Imaging	CapsoVision Inc. (USA)	No	12 to 20	31.0 × 11.3	CapsoView®	CapsoRetrieve®, CapsoAccess® CDAS2
MiroCam™v2	SB Imaging	IntroMedic Co., Ltd. (South Korea)	Yes	3	24.5 × 10.8	MiroView® v2	MiroCam® Receiver
MiroCam™ Navi	SB Imaging and Navigation	IntroMedic Co., Ltd. (South Korea)	N/A	3	24.0 × 11.0	MiroView® v2	MiroCam® Receiver
OMOM® Smart Capsule	SB Imaging	Chongqing Jinshan Science & Technology Co., Ltd (China)	Yes	2	27.9 × 13.0	OMOM® Workstation	OMOM® Image Recorder
OMOM® Controllable Capsule (Magnetic Control)	SB Imaging		N/A	N/A	N/A		OMOM® Image Recorder, handheld steering magnet
EndoCapsule®	SB Imaging	Olympus Corporation (Japan)	Yes	2	26.0 × 11.0	Olympus VE-1 Real Time Viewer	EndoCapsule® Recorder Set
IntelliCap®	Sensory/Drug Delivery	Medimetrics, a Philips Company (Netherlands)	N/A	No Imaging	26.7 × 11.0	N/A	N/A
Enterion™ Capsule	Sensory/Drug Delivery	Quotient Clinical (England)	N/A	No Imaging	32.0 × 11.0	N/A	N/A